

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
 Data File : VN066820.D
 Acq On : 28 Apr 2021 18:15
 Operator : JC/MD
 Sample : PB136019ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136019ZHE#02

Quant Time: Apr 29 04:31:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	327205	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	493874	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	462007	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	193669	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	176935	42.09	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	84.18%		
35) Dibromofluoromethane	7.513	113	154286	50.61	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	101.22%		
50) Toluene-d8	10.032	98	611257	51.08	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	102.16%		
62) 4-Bromofluorobenzene	12.349	95	199891	49.37	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	98.74%		
Target Compounds						
16) Acetone	3.756	43	22190	13.99	ug/l	100
20) Methylene Chloride	4.475	84	16852	4.19	ug/l	93
43) Isopropyl Acetate	8.098	43	28395	3.67	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
Data File : VN066820.D
Acq On : 28 Apr 2021 18:15
Operator : JC/MD
Sample : PB136019ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136019ZHE#02

Quant Time: Apr 29 04:31:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

