

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
 Data File : VN066823.D
 Acq On : 28 Apr 2021 19:28
 Operator : JC/MD
 Sample : PB136019ZHE#05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136019ZHE#05

Quant Time: Apr 29 04:31:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	316851	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	486985	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	449194	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	180427	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	173491	42.62	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 85.24%			
35) Dibromofluoromethane	7.513	113	152545	50.75	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.50%			
50) Toluene-d8	10.029	98	596754	50.57	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 101.14%			
62) 4-Bromofluorobenzene	12.349	95	186793	46.79	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 93.58%			
Target Compounds					Qvalue	
16) Acetone	3.756	43	22333	14.54	ug/l	98
20) Methylene Chloride	4.475	84	29774	7.65	ug/l	98
43) Isopropyl Acetate	8.095	43	26269	3.44	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
Data File : VN066823.D
Acq On : 28 Apr 2021 19:28
Operator : JC/MD
Sample : PB136019ZHE#05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136019ZHE#05

Quant Time: Apr 29 04:31:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

