

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
 Data File : VN066824.D
 Acq On : 28 Apr 2021 19:52
 Operator : JC/MD
 Sample : M2138-13
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-MW-15D

Quant Time: Apr 29 04:32:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	303405	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	464804	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	408830	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	169794	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	167158	42.88	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	85.76%
35) Dibromofluoromethane	7.513	113	146453	51.04	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.08%
50) Toluene-d8	10.032	98	556530	49.42	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.84%
62) 4-Bromofluorobenzene	12.351	95	148265	38.91	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	77.82%
Target Compounds						
						Qvalue
30) Chloroform	7.293	83	3475	0.57	ug/l	97
64) Tetrachloroethene	10.573	164	137774	50.96	ug/l	98
90) Hexachloroethane	13.816	117	4011	1.91	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
Data File : VN066824.D
Acq On : 28 Apr 2021 19:52
Operator : JC/MD
Sample : M2138-13
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-MW-15D

Quant Time: Apr 29 04:32:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

