

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
 Data File : VN066833.D
 Acq On : 29 Apr 2021 00:29
 Operator : JC/MD
 Sample : M2137-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-DDC-6-PS

Quant Time: Apr 29 04:33:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	315598	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	477505	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	418494	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	167280	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	170844	42.13	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 84.26%			
35) Dibromofluoromethane	7.516	113	150453	51.04	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 102.08%			
50) Toluene-d8	10.029	98	571550	49.40	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 98.80%			
62) 4-Bromofluorobenzene	12.349	95	164799	42.10	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 84.20%			
Target Compounds						
27) cis-1,2-Dichloroethene	6.744	96	19826	5.04	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
Data File : VN066833.D
Acq On : 29 Apr 2021 00:29
Operator : JC/MD
Sample : M2137-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-6-PS

Quant Time: Apr 29 04:33:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

