

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
 Data File : VN066834.D
 Acq On : 29 Apr 2021 00:53
 Operator : JC/MD
 Sample : M2137-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-DDC-6-PD

Quant Time: Apr 29 04:34:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	299998	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	445440	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	380226	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	146375	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	161955	42.02	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	84.04%		
35) Dibromofluoromethane	7.516	113	143649	52.24	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	104.48%		
50) Toluene-d8	10.032	98	536112	49.67	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	99.34%		
62) 4-Bromofluorobenzene	12.351	95	123646	33.86	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	67.72%		
Target Compounds						
27) cis-1,2-Dichloroethene	6.738	96	17139	4.59	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
Data File : VN066834.D
Acq On : 29 Apr 2021 00:53
Operator : JC/MD
Sample : M2137-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-6-PD

Quant Time: Apr 29 04:34:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

