

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
 Data File : VN066835.D
 Acq On : 29 Apr 2021 1:17
 Operator : JC/MD
 Sample : M2137-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-MW-3D(OFFSITE)

Quant Time: Apr 29 04:34:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.594	168	318157	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	478439	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	399596	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	170324	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	171415	41.93	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 83.86%			
35) Dibromofluoromethane	7.516	113	153338	51.92	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 103.84%			
50) Toluene-d8	10.032	98	568510	49.04	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 98.08%			
62) 4-Bromofluorobenzene	12.351	95	160156	40.83	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 81.66%			
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	6.743	96	53679	13.55	ug/l	90
30) Chloroform	7.293	83	1766	0.28	ug/l	92
44) Trichloroethene	8.771	130	12236	3.64	ug/l	96
64) Tetrachloroethene	10.573	164	19588	7.41	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
Data File : VN066835.D
Acq On : 29 Apr 2021 1:17
Operator : JC/MD
Sample : M2137-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-MW-3D(OFFSITE)

Quant Time: Apr 29 04:34:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

