

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
 Data File : VN066840.D
 Acq On : 29 Apr 2021 3:18
 Operator : JC/MD
 Sample : M2137-09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-DDC-10-PS

Quant Time: Apr 29 04:34:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	303094	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	460765	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	392803	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	155257	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	165372	42.47	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	84.94%		
35) Dibromofluoromethane	7.516	113	148839	52.33	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	104.66%		
50) Toluene-d8	10.031	98	547404	49.03	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.06%		
62) 4-Bromofluorobenzene	12.351	95	149634	39.61	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	79.22%		
Target Compounds						
27) cis-1,2-Dichloroethene	6.741	96	11059	2.93	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
Data File : VN066840.D
Acq On : 29 Apr 2021 3:18
Operator : JC/MD
Sample : M2137-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-10-PS

Quant Time: Apr 29 04:34:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

