

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
 Data File : VN066844.D
 Acq On : 29 Apr 2021 4:53
 Operator : JC/MD
 Sample : M2137-13
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-DDC-7-PS

Quant Time: Apr 29 07:45:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.594	168	296431	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	446367	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	385928	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	145354	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	164034	43.07	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	86.14%
35) Dibromofluoromethane	7.516	113	145856	52.94	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.88%
50) Toluene-d8	10.032	98	541812	50.10	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.20%
62) 4-Bromofluorobenzene	12.351	95	125082	34.18	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	68.36%
Target Compounds						
					Qvalue	
16) Acetone	3.766	43	3742	2.60	ug/l	96
27) cis-1,2-Dichloroethene	6.743	96	14839	4.02	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
Data File : VN066844.D
Acq On : 29 Apr 2021 4:53
Operator : JC/MD
Sample : M2137-13
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-7-PS

Quant Time: Apr 29 07:45:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

