

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
 Data File : VN066845.D
 Acq On : 29 Apr 2021 5:17
 Operator : JC/MD
 Sample : M2170-18
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AN-1

Manual Integrations
 APPROVED

MMDadoda
 4/29/2021 5:52:22 PM

Quant Time: Apr 29 07:46:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	320356	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	486474	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	459485	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	187806	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	175233	42.57	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	85.14%		
35) Dibromofluoromethane	7.513	113	157116	52.32	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	104.64%		
50) Toluene-d8	10.029	98	610350	51.78	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	103.56%		
62) 4-Bromofluorobenzene	12.349	95	181278	45.45	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	90.90%		
Target Compounds						Qvalue
16) Acetone	3.759	43	31873	20.53	ug/l	96
18) Methyl Acetate	4.260	43	3021m	1.82	ug/l	
20) Methylene Chloride	4.475	84	21496	5.46	ug/l	95
43) Isopropyl Acetate	8.095	43	23984	3.14	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
Data File : VN066845.D
Acq On : 29 Apr 2021 5:17
Operator : JC/MD
Sample : M2170-18
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AN-1

Manual Integrations
APPROVED

MMDadoda
4/29/2021 5:52:22 PM

Quant Time: Apr 29 07:46:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

