

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
 Data File : VN066847.D
 Acq On : 29 Apr 2021 6:05
 Operator : JC/MD
 Sample : M2187-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TEST-PIT-II

Quant Time: Apr 29 07:46:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.594	168	304164	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	464794	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	445834	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	184187	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	167962	42.98	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	85.96%
35) Dibromofluoromethane	7.513	113	150684	52.52	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.04%
50) Toluene-d8	10.029	98	585424	51.98	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.96%
62) 4-Bromofluorobenzene	12.349	95	175750	46.12	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.24%
Target Compounds						
16) Acetone	3.756	43	85545	58.02	ug/l	95
18) Methyl Acetate	4.263	43	3872	2.07	ug/l #	65
20) Methylene Chloride	4.475	84	12485	3.34	ug/l	93
43) Isopropyl Acetate	8.098	43	23830	3.27	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
Data File : VN066847.D
Acq On : 29 Apr 2021 6:05
Operator : JC/MD
Sample : M2187-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TEST-PIT-II

Quant Time: Apr 29 07:46:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

