

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
 Data File : VN066848.D
 Acq On : 29 Apr 2021 6:29
 Operator : JC/MD
 Sample : M2196-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WALL-D-4075-A

Quant Time: Apr 29 07:47:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.594	168	306432	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	458219	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	428681	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	156190	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	168298	42.75	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	85.50%
35) Dibromofluoromethane	7.513	113	150154	53.09	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.18%
50) Toluene-d8	10.032	98	580702	52.30	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.60%
62) 4-Bromofluorobenzene	12.349	95	163329	43.48	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.96%
Target Compounds						
16) Acetone	3.756	43	30116	20.28	ug/l	94
18) Methyl Acetate	4.263	43	4005	2.09	ug/l #	75
20) Methylene Chloride	4.475	84	13678	3.63	ug/l	93
43) Isopropyl Acetate	8.095	43	24592	3.42	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
Data File : VN066848.D
Acq On : 29 Apr 2021 6:29
Operator : JC/MD
Sample : M2196-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WALL-D-4075-A

Quant Time: Apr 29 07:47:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

