


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN042821\  
Data File : VN066827.D  
Acq On : 28 Apr 2021 21:28  
Operator : JC/MD  
Sample : BFB  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 30 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N042721W.M  
Title : SW846 8260  
Last Update : Tue Apr 27 16:02:50 2021



AutoFind: Scans 3962, 3963, 3964; Background Corrected with Scan 3946

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 16.0         | 6043       | PASS                |
| 75             | 95              | 30              | 60              | 45.2         | 17051      | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 37704      | PASS                |
| 96             | 95              | 5               | 9               | 6.6          | 2495       | PASS                |
| 173            | 174             | 0.00            | 2               | 1.0          | 322        | PASS                |
| 174            | 95              | 50              | 100             | 85.6         | 32275      | PASS                |
| 175            | 174             | 5               | 9               | 7.5          | 2412       | PASS                |
| 176            | 174             | 95              | 101             | 98.4         | 31747      | PASS                |
| 177            | 176             | 5               | 9               | 6.9          | 2194       | PASS                |