

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042822\
 Data File : VN072239.D
 Acq On : 28 Apr 2022 15:41
 Operator : JC\MD
 Sample : N2601-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-MW150S1-GW-235-237

Quant Time: Apr 29 06:33:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	247664	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	460133	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	484086	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	172818	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	193000	50.760	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.520%
35) Dibromofluoromethane	8.016	113	141852	48.165	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.340%
50) Toluene-d8	10.439	98	587109	50.733	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.460%
62) 4-Bromofluorobenzene	12.727	95	225442	58.691	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	117.380%
Target Compounds						
						Qvalue
16) Acetone	4.292	43	5272	3.216	ug/l	95
30) Chloroform	7.816	83	1722	0.303	ug/l	85
44) Trichloroethene	9.210	130	33035	10.680	ug/l	93
64) Tetrachloroethene	10.963	164	2887	0.991	ug/l #	60

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042822\
Data File : VN072239.D
Acq On : 28 Apr 2022 15:41
Operator : JC\MD
Sample : N2601-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-MW150S1-GW-235-237

Quant Time: Apr 29 06:33:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

