

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042822\
 Data File : VN072240.D
 Acq On : 28 Apr 2022 16:05
 Operator : JC\MD
 Sample : N2601-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-MW150S1-GW-245-247

Quant Time: Apr 29 06:34:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	240886	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	451190	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	498641	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	183367	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	196548	53.148	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 106.300%			
35) Dibromofluoromethane	8.022	113	140596	48.685	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 97.380%			
50) Toluene-d8	10.439	98	585119	51.563	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 103.120%			
62) 4-Bromofluorobenzene	12.727	95	232440	61.713	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 123.420%			
Target Compounds					Qvalue	
16) Acetone	4.304	43	5790	3.631	ug/l	94
25) 2-Butanone	7.351	43	1606	0.632	ug/l	95
30) Chloroform	7.816	83	2135	0.386	ug/l #	72
44) Trichloroethene	9.210	130	65387	21.558	ug/l	89
64) Tetrachloroethene	10.974	164	2677	0.892	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042822\
Data File : VN072240.D
Acq On : 28 Apr 2022 16:05
Operator : JC\MD
Sample : N2601-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-MW150S1-GW-245-247

Quant Time: Apr 29 06:34:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

