

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042822\
 Data File : VN072247.D
 Acq On : 28 Apr 2022 18:59
 Operator : JC\MD
 Sample : N2617-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-13

Quant Time: Apr 29 06:36:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	258813	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	482193	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	534702	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	196984	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	204801	51.544	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.080%
35) Dibromofluoromethane	8.022	113	148475	48.108	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.220%
50) Toluene-d8	10.439	98	621396	51.239	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.480%
62) 4-Bromofluorobenzene	12.727	95	254293	63.174	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	126.340%
Target Compounds						
						Qvalue
19) Methyl tert-butyl Ether	5.628	73	2989	0.308	ug/l	# 89
67) Ethyl Benzene	11.839	91	10987	0.583	ug/l	95
68) m/p-Xylenes	11.939	106	3295	0.471	ug/l	88
95) Naphthalene	15.504	128	17719	5.887	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042822\
Data File : VN072247.D
Acq On : 28 Apr 2022 18:59
Operator : JC\MD
Sample : N2617-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-13

Quant Time: Apr 29 06:36:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

