

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042822\
 Data File : VN072254.D
 Acq On : 28 Apr 2022 21:49
 Operator : JC\MD
 Sample : N2617-17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-18I

Quant Time: Apr 29 06:38:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	273378	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	504755	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	547222	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	196407	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	211066	50.290	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 100.580%		
35) Dibromofluoromethane	8.016	113	154411	47.795	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 95.580%		
50) Toluene-d8	10.439	98	647996	51.044	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 102.080%		
62) 4-Bromofluorobenzene	12.727	95	252431	59.908	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 119.820%		
Target Compounds						
					Qvalue	
30) Chloroform	7.822	83	6664	1.063	ug/l #	77
47) Bromodichloromethane	9.751	83	2035	0.393	ug/l #	94
67) Ethyl Benzene	11.839	91	6904	0.358	ug/l #	88
68) m/p-Xylenes	11.945	106	3089	0.431	ug/l #	64
95) Naphthalene	15.504	128	7474	4.971	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042822\
Data File : VN072254.D
Acq On : 28 Apr 2022 21:49
Operator : JC\MD
Sample : N2617-17
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-18I

Quant Time: Apr 29 06:38:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

