

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042822\
 Data File : VN072258.D
 Acq On : 28 Apr 2022 23:25
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 29 06:39:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.621	0.520	16.3	89	0.00
3 P	Chloromethane	0.762	0.540	29.1#	82	0.00
4 C	Vinyl Chloride	0.563	0.478	15.1#	87	0.00
5 T	Bromomethane	0.307	0.265	13.7	92	0.00
6 T	Chloroethane	0.372	0.286	23.1	75	0.00
7 T	Trichlorofluoromethane	0.983	0.881	10.4	92	0.00
8 T	Diethyl Ether	0.356	0.332	6.7	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.517	0.453	12.4	87	0.00
10 T	Methyl Iodide	0.627	0.578	7.8	87	0.00
11 T	Tert butyl alcohol	0.126	0.131	-4.0	98	0.00
12 CM	1,1-Dichloroethene	0.478	0.415	13.2#	84	0.00
13 T	Acrolein	0.127	0.110	13.4	84	0.00
14 T	Allyl chloride	0.992	0.850	14.3	82	0.00
15 T	Acrylonitrile	0.381	0.352	7.6	86	0.00
16 T	Acetone	0.331	0.301	9.1	90	0.00
17 T	Carbon Disulfide	1.222	0.956	21.8	82	0.00
18 T	Methyl Acetate	1.039	0.827	20.4	85	0.00
19 T	Methyl tert-butyl Ether	1.872	1.875	-0.2	93	0.00
20 T	Methylene Chloride	0.634	0.541	14.7	88	0.01
21 T	trans-1,2-Dichloroethene	0.527	0.462	12.3	87	0.00
22 T	Diisopropyl ether	2.042	1.829	10.4	87	0.00
23 T	Vinyl Acetate	1.715	1.587	7.5	86	0.00
24 P	1,1-Dichloroethane	1.142	1.018	10.9	87	0.00
25 T	2-Butanone	0.527	0.474	10.1	86	0.00
26 T	2,2-Dichloropropane	0.892	0.724	18.8	78	0.00
27 T	cis-1,2-Dichloroethene	0.642	0.575	10.4	88	0.00
28 T	Bromochloromethane	0.512	0.435	15.0	79	0.00
29 T	Tetrahydrofuran	0.346	0.309	10.7	84	0.00
30 C	Chloroform	1.147	1.046	8.8#	92	0.00
31 T	Cyclohexane	1.026	0.830	19.1	83	0.00
32 T	1,1,1-Trichloroethane	0.976	0.945	3.2	94	0.00
33 S	1,2-Dichloroethane-d4	0.768	0.758	1.3	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.320	0.303	5.3	91	0.00
36 T	1,1-Dichloropropene	0.466	0.417	10.5	87	0.00
37 T	Ethyl Acetate	0.619	0.546	11.8	87	0.00
38 T	Carbon Tetrachloride	0.499	0.465	6.8	95	0.00
39 T	Methylcyclohexane	0.521	0.471	9.6	85	0.00
40 TM	Benzene	1.445	1.272	12.0	88	0.00
41 T	Methacrylonitrile	0.313	0.287	8.3	85	0.00
42 TM	1,2-Dichloroethane	0.553	0.510	7.8	92	0.00
43 T	Isopropyl Acetate	1.015	0.842	17.0	87	0.00
44 TM	Trichloroethene	0.336	0.292	13.1	88	0.00
45 C	1,2-Dichloropropane	0.390	0.335	14.1#	85	0.00
46 T	Dibromomethane	0.245	0.224	8.6	90	0.00
47 T	Bromodichloromethane	0.513	0.488	4.9	91	0.00
48 T	Methyl methacrylate	0.409	0.382	6.6	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042822\
 Data File : VN072258.D
 Acq On : 28 Apr 2022 23:25
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 29 06:39:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.007	0.0	89	0.00
50 S	Toluene-d8	1.258	1.205	4.2	90	0.00
51 T	4-Methyl-2-Pentanone	0.584	0.546	6.5	85	0.00
52 CM	Toluene	0.864	0.812	6.0#	88	0.00
53 T	t-1,3-Dichloropropene	0.516	0.503	2.5	88	0.00
54 T	cis-1,3-Dichloropropene	0.558	0.534	4.3	87	0.00
55 T	1,1,2-Trichloroethane	0.362	0.323	10.8	87	0.00
56 T	Ethyl methacrylate	0.537	0.565	-5.2	90	0.00
57 T	1,3-Dichloropropane	0.636	0.583	8.3	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.129	0.146	-13.2	88	0.00
59 T	2-Hexanone	0.411	0.396	3.6	85	0.00
60 T	Dibromochloromethane	0.370	0.359	3.0	91	0.00
61 T	1,2-Dibromoethane	0.356	0.332	6.7	89	0.00
62 S	4-Bromofluorobenzene	0.417	0.425	-1.9	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.301	0.270	10.3	89	0.00
65 PM	Chlorobenzene	1.010	0.908	10.1	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.381	0.361	5.2	90	0.00
67 C	Ethyl Benzene	1.762	1.767	-0.3#	91	0.00
68 T	m/p-Xylenes	0.654	0.643	1.7	89	0.00
69 T	o-Xylene	0.634	0.631	0.5	87	0.00
70 T	Styrene	0.999	1.023	-2.4	88	0.00
71 P	Bromoform	0.283	0.282	0.4	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	4.416	4.242	3.9	90	0.00
74 T	N-amyl acetate	1.654	1.557	5.9	85	0.00
75 P	1,1,2,2-Tetrachloroethane	1.719	1.391	19.1	87	0.00
76 T	1,2,3-Trichloropropane	1.445	1.113	23.0	90	0.00
77 T	Bromobenzene	1.044	0.922	11.7	90	0.00
78 T	n-propylbenzene	4.865	4.629	4.9	87	0.00
79 T	2-Chlorotoluene	3.195	2.938	8.0	89	0.00
80 T	1,3,5-Trimethylbenzene	3.678	3.528	4.1	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.432	0.390	9.7	81	0.00
82 T	4-Chlorotoluene	2.975	2.795	6.1	88	0.00
83 T	tert-Butylbenzene	3.229	3.087	4.4	89	0.00
84 T	1,2,4-Trimethylbenzene	3.558	3.518	1.1	89	0.00
85 T	sec-Butylbenzene	4.210	4.083	3.0	88	0.00
86 T	p-Isopropyltoluene	3.282	3.351	-2.1	88	0.00
87 T	1,3-Dichlorobenzene	1.690	1.554	8.0	87	0.00
88 T	1,4-Dichlorobenzene	1.644	1.504	8.5	87	0.00
89 T	n-Butylbenzene	2.514	2.511	0.1	84	0.00
90 T	Hexachloroethane	0.712	0.616	13.5	88	0.00
91 T	1,2-Dichlorobenzene	1.666	1.520	8.8	88	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.305	0.281	7.9	88	0.00
93 T	1,2,4-Trichlorobenzene	0.656	0.695	-5.9	87	0.00
94 T	Hexachlorobutadiene	0.446	0.400	10.3	90	0.00
95 T	Naphthalene	1.870	2.439	-30.4#	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042822\
Data File : VN072258.D
Acq On : 28 Apr 2022 23:25
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Apr 29 06:39:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.634	0.719	-13.4	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6