

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042825\
 Data File : VN086400.D
 Acq On : 28 Apr 2025 10:34
 Operator : JC\MD
 Sample : VN0428MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0428MBL01

Quant Time: Apr 29 01:47:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	178399	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	334831	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	302710	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	127091	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	124722	48.212	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.420%
35) Dibromofluoromethane	8.165	113	88614	57.023	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	114.040%
50) Toluene-d8	10.565	98	415458	50.023	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.040%
62) 4-Bromofluorobenzene	12.847	95	143240	47.285	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042825\
Data File : VN086400.D
Acq On : 28 Apr 2025 10:34
Operator : JC\MD
Sample : VN0428MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0428MBL01

Quant Time: Apr 29 01:47:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

