

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042825\
 Data File : VN086401.D
 Acq On : 28 Apr 2025 10:58
 Operator : JC\MD
 Sample : VN0428WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0428WBL01

Quant Time: Apr 29 01:48:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	202253	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	381951	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	370420	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	158007	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	139828	47.677	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.360%
35) Dibromofluoromethane	8.165	113	99415	56.082	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.160%
50) Toluene-d8	10.565	98	492452	51.979	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.960%
62) 4-Bromofluorobenzene	12.847	95	183407	53.075	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042825\
Data File : VN086401.D
Acq On : 28 Apr 2025 10:58
Operator : JC\MD
Sample : VN0428WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0428WBL01

Quant Time: Apr 29 01:48:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

