

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042825\
 Data File : VN086405.D
 Acq On : 28 Apr 2025 12:44
 Operator : JC\MD
 Sample : PB167744ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167744ZHE#02

Quant Time: Apr 29 01:51:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	180833	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	341859	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	319771	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	143972	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	127174	48.498	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.000%
35) Dibromofluoromethane	8.165	113	84440	53.220	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.440%
50) Toluene-d8	10.565	98	435655	51.377	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.760%
62) 4-Bromofluorobenzene	12.847	95	156440	50.581	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.160%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	65178	67.731	ug/l	96
20) Methylene Chloride	5.265	84	3929	1.621	ug/l #	85
43) Isopropyl Acetate	8.682	43	24070	2.200	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042825\
Data File : VN086405.D
Acq On : 28 Apr 2025 12:44
Operator : JC\MD
Sample : PB167744ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167744ZHE#02

Quant Time: Apr 29 01:51:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

