

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042825\
 Data File : VN086410.D
 Acq On : 28 Apr 2025 14:45
 Operator : JC\MD
 Sample : PB167744ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167744ZHE#07

Quant Time: Apr 29 01:53:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	173616	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	337344	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	325417	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	141094	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	126074	50.077	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 100.160%		
35) Dibromofluoromethane	8.165	113	83365	53.246	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.500%		
50) Toluene-d8	10.565	98	430762	51.480	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.960%		
62) 4-Bromofluorobenzene	12.847	95	159184	52.157	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 104.320%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	62611	67.771	ug/l	96
20) Methylene Chloride	5.277	84	4319	1.856	ug/l	96
43) Isopropyl Acetate	8.682	43	29770	3.234	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042825\
Data File : VN086410.D
Acq On : 28 Apr 2025 14:45
Operator : JC\MD
Sample : PB167744ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167744ZHE#07

Quant Time: Apr 29 01:53:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

