

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042825\
 Data File : VN086418.D
 Acq On : 29 Apr 2025 01:44
 Operator : JC\MD
 Sample : PB167744ZHE#15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167744ZHE#15

Quant Time: Apr 29 03:40:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	60784	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	129955	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.859	117	119374	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	48448	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	76677	86.992	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	173.980%#
35) Dibromofluoromethane	8.159	113	41563	68.911	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	137.820%#
50) Toluene-d8	10.559	98	158092	49.044	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.080%
62) 4-Bromofluorobenzene	12.841	95	56051	47.673	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.340%
Target Compounds						
						Qvalue
16) Acetone	4.442	43	47639	154.136	ug/l	93
18) Methyl Acetate	5.036	43	4747	4.487	ug/l #	86
20) Methylene Chloride	5.271	84	1589	1.950	ug/l #	80
25) 2-Butanone	7.489	43	5531	10.081	ug/l #	76
43) Isopropyl Acetate	8.688	43	40217	16.159	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042825\
Data File : VN086418.D
Acq On : 29 Apr 2025 01:44
Operator : JC\MD
Sample : PB167744ZHE#15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167744ZHE#15

Quant Time: Apr 29 03:40:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

