

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042825\
 Data File : VN086422.D
 Acq On : 29 Apr 2025 03:20
 Operator : JC\MD
 Sample : PB167744TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167744TB

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/29/2025
 Supervised By : Semsettin Yesilyurt 04/29/2025

Quant Time: Apr 29 03:42:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	176402	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	339849	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	323035	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	141231	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	128057	50.062	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.120%			
35) Dibromofluoromethane	8.165	113	86856	55.067	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 110.140%			
50) Toluene-d8	10.565	98	431508	51.189	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.380%			
62) 4-Bromofluorobenzene	12.847	95	158800	51.647	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 103.300%			
Target Compounds					Qvalue	
16) Acetone	4.430	43	60572	64.249	ug/l	94
18) Methyl Acetate	5.030	43	3707	1.207	ug/l #	69
20) Methylene Chloride	5.277	84	2675	1.131	ug/l	89
43) Isopropyl Acetate	8.682	43	34675m	4.034	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042825\
Data File : VN086422.D
Acq On : 29 Apr 2025 03:20
Operator : JC\MD
Sample : PB167744TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167744TB

Quant Time: Apr 29 03:42:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/29/2025
Supervised By :Semsettin Yesilyurt 04/29/2025

