

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042919\
 Data File : VN055291.D
 Acq On : 29 Apr 2019 14:13
 Operator : JC/SP
 Sample : PB119269TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB119269TB

Quant Time: Apr 30 04:55:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 04:20:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	398540	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	669240	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	632393	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	179843	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	272080	52.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.36%	
35) Dibromofluoromethane	7.59	113	223746	50.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.88%	
50) Toluene-d8	10.09	98	851046	53.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.32%	
62) 4-Bromofluorobenzene	12.40	95	266997	53.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.72%	
Target Compounds						
16) Acetone	3.82	43	11591	7.135	ug/l	95
20) Methylene Chloride	4.55	84	4466	0.901	ug/l #	90
43) Isopropyl Acetate	8.16	43	18040	2.285	ug/l #	89
80) 1,3,5-Trimethylbenzene	12.73	105	13303	0.944	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042919\
Data File : VN055291.D
Acq On : 29 Apr 2019 14:13
Operator : JC/SP
Sample : PB119269TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB119269TB

Quant Time: Apr 30 04:55:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 04:20:38 2019
Response via : Initial Calibration

