

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042920\
 Data File : VN061188.D
 Acq On : 29 Apr 2020 13:27
 Operator : JC/MD
 Sample : L2415-07DL 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ST008RW256-200423DL

Quant Time: Apr 30 14:09:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 29 08:14:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	94245	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	166026	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	162451	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	73211	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	75621	50.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.42%	
35) Dibromofluoromethane	7.55	113	51348	52.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.60%	
50) Toluene-d8	10.06	98	207763	51.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.82%	
62) 4-Bromofluorobenzene	12.37	95	78663	52.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.46%	
Target Compounds						
16) Acetone	3.78	43	3386	0.354	ug/l	94
27) cis-1,2-Dichloroethene	6.78	96	19427	15.289	ug/l	87
44) Trichloroethene	8.80	130	11935	8.916	ug/l	94
64) Tetrachloroethene	10.60	164	24879	17.682	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042920\
Data File : VN061188.D
Acq On : 29 Apr 2020 13:27
Operator : JC/MD
Sample : L2415-07DL 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ST008RW256-200423DL

Quant Time: Apr 30 14:09:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 29 08:14:26 2020
Response via : Initial Calibration

