

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042920\
 Data File : VN061196.D
 Acq On : 29 Apr 2020 16:56
 Operator : JC/MD
 Sample : PB128535TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB128535TB

Quant Time: Apr 30 03:37:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 29 08:14:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	92304	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	167439	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	171770	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	75815	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	76626	52.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.92%	
35) Dibromofluoromethane	7.55	113	47170	47.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.28%	
50) Toluene-d8	10.06	98	217619	53.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.84%	
62) 4-Bromofluorobenzene	12.38	95	82036	54.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.02%	
Target Compounds						
16) Acetone	3.77	43	11348	14.894	ug/l	97
43) Isopropyl Acetate	8.13	43	36108	11.072	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042920\
Data File : VN061196.D
Acq On : 29 Apr 2020 16:56
Operator : JC/MD
Sample : PB128535TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB128535TB

Quant Time: Apr 30 03:37:29 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 29 08:14:26 2020
Response via : Initial Calibration

