

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
 Data File : VN066854.D
 Acq On : 29 Apr 2021 10:35
 Operator : JC/MD
 Sample : VN0429MBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0429MBL01

Quant Time: Apr 30 06:13:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.588	168	309156	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	470627	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	394728	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	179772	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.950	65	166871	42.01	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	84.02%
35) Dibromofluoromethane	7.511	113	152613	52.53	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.06%
50) Toluene-d8	10.029	98	560556	49.16	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.32%
62) 4-Bromofluorobenzene	12.349	95	165746	42.96	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	85.92%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
Data File : VN066854.D
Acq On : 29 Apr 2021 10:35
Operator : JC/MD
Sample : VN0429MBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0429MBL01

Quant Time: Apr 30 06:13:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

