

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
 Data File : VN066858.D
 Acq On : 29 Apr 2021 13:51
 Operator : JC/MD
 Sample : M2173-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Apr 30 06:14:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.594	168	304536	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.521	114	455092	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	389891	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	150923	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	174647	44.63	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	89.26%
35) Dibromofluoromethane	7.516	113	159089	56.63	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	113.26%
50) Toluene-d8	10.031	98	590178	53.52	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.04%
62) 4-Bromofluorobenzene	12.351	95	164937	44.21	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.42%
Target Compounds						
95) Naphthalene	15.092	128	21094	3.15	ug/l	Qvalue # 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
Data File : VN066858.D
Acq On : 29 Apr 2021 13:51
Operator : JC/MD
Sample : M2173-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Apr 30 06:14:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

