

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
 Data File : VN066859.D
 Acq On : 29 Apr 2021 14:15
 Operator : JC/MD
 Sample : M2173-02RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5R

Quant Time: Apr 30 06:14:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	321007	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	478788	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	414630	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	156388	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	183061	44.38	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	88.76%
35) Dibromofluoromethane	7.513	113	165091	55.86	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	111.72%
50) Toluene-d8	10.029	98	615214	53.03	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.06%
62) 4-Bromofluorobenzene	12.349	95	147757	37.64	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	75.28%
Target Compounds						
38) Carbon Tetrachloride	7.591	117	26613	6.24	ug/l	Qvalue # 18

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
Data File : VN066859.D
Acq On : 29 Apr 2021 14:15
Operator : JC/MD
Sample : M2173-02RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-5R

Quant Time: Apr 30 06:14:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

