

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
 Data File : VN066861.D
 Acq On : 29 Apr 2021 14:58
 Operator : JC/MD
 Sample : M2173-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Apr 30 06:15:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.596	168	309668	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.522	114	470272	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	397680	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	159369	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	181237	45.55	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.10%
35) Dibromofluoromethane	7.519	113	162620	56.02	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	112.04%
50) Toluene-d8	10.032	98	603644	52.98	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.96%
62) 4-Bromofluorobenzene	12.352	95	166312	43.14	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.28%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
Data File : VN066861.D
Acq On : 29 Apr 2021 14:58
Operator : JC/MD
Sample : M2173-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Apr 30 06:15:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

