

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
 Data File : VN066862.D
 Acq On : 29 Apr 2021 15:19
 Operator : JC/MD
 Sample : M2138-17
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-DDC-4-PS

Quant Time: Apr 30 06:15:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.596	168	304775	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.522	114	462269	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.354	117	391963	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	162166	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.958	65	178206	45.51	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.02%
35) Dibromofluoromethane	7.516	113	158293	55.47	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	110.94%
50) Toluene-d8	10.032	98	594181	53.05	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.10%
62) 4-Bromofluorobenzene	12.351	95	165640	43.71	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.42%
Target Compounds						
64) Tetrachloroethene	10.576	164	11634	4.49	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
Data File : VN066862.D
Acq On : 29 Apr 2021 15:19
Operator : JC/MD
Sample : M2138-17
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-4-PS

Quant Time: Apr 30 06:15:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

