

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
 Data File : VN066864.D
 Acq On : 29 Apr 2021 16:03
 Operator : JC/MD
 Sample : M2138-13
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-MW-15D

Quant Time: Apr 30 06:15:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.594	168	305798	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.522	114	461773	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	409035	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	162868	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	178192	45.35	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 90.70%			
35) Dibromofluoromethane	7.513	113	160188	56.20	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 112.40%			
50) Toluene-d8	10.031	98	589345	52.67	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 105.34%			
62) 4-Bromofluorobenzene	12.351	95	167717	44.30	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 88.60%			
Target Compounds						
30) Chloroform	7.291	83	3390m	0.55	ug/l	Qvalue
64) Tetrachloroethene	10.573	164	142085	52.53	ug/l	98
90) Hexachloroethane	13.816	117	3879	1.93	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
Data File : VN066864.D
Acq On : 29 Apr 2021 16:03
Operator : JC/MD
Sample : M2138-13
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-MW-15D

Quant Time: Apr 30 06:15:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

