

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
 Data File : VN066869.D
 Acq On : 29 Apr 2021 18:03
 Operator : JC/MD
 Sample : M2137-05RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-DDC-5-PSRE

Quant Time: Apr 30 06:16:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.594	168	345061	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	525873	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	448180	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	185687	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	199195	44.93	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 89.86%			
35) Dibromofluoromethane	7.513	113	179338	55.25	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 110.50%			
50) Toluene-d8	10.031	98	657009	51.56	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 103.12%			
62) 4-Bromofluorobenzene	12.351	95	160436	37.21	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 74.42%			
Target Compounds						
27) cis-1,2-Dichloroethene	6.743	96	14638	3.41	ug/l	92
64) Tetrachloroethene	10.573	164	3387	1.14	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
Data File : VN066869.D
Acq On : 29 Apr 2021 18:03
Operator : JC/MD
Sample : M2137-05RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-5-PSRE

Quant Time: Apr 30 06:16:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

