

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
 Data File : VN066873.D
 Acq On : 29 Apr 2021 19:40
 Operator : JC/MD
 Sample : M2137-10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-MW-2D(OFFSITE)

Quant Time: Apr 30 06:16:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	300079	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.522	114	455592	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	398220	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	153965	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	178412	46.27	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.54%
35) Dibromofluoromethane	7.513	113	160944	57.23	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	114.46%
50) Toluene-d8	10.032	98	586526	53.13	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.26%
62) 4-Bromofluorobenzene	12.352	95	160189	42.89	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	85.78%
Target Compounds						
27) cis-1,2-Dichloroethene	6.744	96	11063	2.96	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
Data File : VN066873.D
Acq On : 29 Apr 2021 19:40
Operator : JC/MD
Sample : M2137-10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-MW-2D(OFFSITE)

Quant Time: Apr 30 06:16:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

