

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
 Data File : VN066874.D
 Acq On : 29 Apr 2021 20:04
 Operator : JC/MD
 Sample : M2137-12
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-DDC-8-PS

Quant Time: Apr 30 06:16:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	303497	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	457112	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	405387	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	154725	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	179800	46.11	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	92.22%		
35) Dibromofluoromethane	7.513	113	163068	57.79	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	115.58%		
50) Toluene-d8	10.032	98	592596	53.50	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	107.00%		
62) 4-Bromofluorobenzene	12.352	95	165177	44.08	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	88.16%		
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	6.746	96	12632	3.34	ug/l	88
30) Chloroform	7.291	83	1785	0.29	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
Data File : VN066874.D
Acq On : 29 Apr 2021 20:04
Operator : JC/MD
Sample : M2137-12
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-8-PS

Quant Time: Apr 30 06:16:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

