

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
 Data File : VN066876.D
 Acq On : 29 Apr 2021 20:52
 Operator : JC/MD
 Sample : M2137-14
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-DDC-9-PD

Quant Time: Apr 30 06:17:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	357281	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.522	114	546072	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	454096	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	179959	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	212040	46.19	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.38%
35) Dibromofluoromethane	7.513	113	190722	56.58	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	113.16%
50) Toluene-d8	10.032	98	692695	52.35	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.70%
62) 4-Bromofluorobenzene	12.349	95	187365	41.85	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	83.70%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
Data File : VN066876.D
Acq On : 29 Apr 2021 20:52
Operator : JC/MD
Sample : M2137-14
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-9-PD

Quant Time: Apr 30 06:17:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

