

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
 Data File : VN066884.D
 Acq On : 30 Apr 2021 00:52
 Operator : JC/MD
 Sample : M2137-15
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-DDC-7-PD

Quant Time: Apr 30 06:18:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	302476	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	453861	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	394457	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	157838	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	177515	45.68	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.36%
35) Dibromofluoromethane	7.516	113	161529	57.66	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	115.32%
50) Toluene-d8	10.032	98	588302	53.50	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.00%
62) 4-Bromofluorobenzene	12.349	95	165103	44.37	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.74%
Target Compounds						
27) cis-1,2-Dichloroethene	6.743	96	6281	1.67	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
Data File : VN066884.D
Acq On : 30 Apr 2021 00:52
Operator : JC/MD
Sample : M2137-15
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-7-PD

Quant Time: Apr 30 06:18:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

