

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
 Data File : VN066885.D
 Acq On : 30 Apr 2021 1:17
 Operator : JC/MD
 Sample : M2137-16
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-DDC-9-PS

Manual Integrations
 APPROVED

MMDadoda
 4/30/2021 11:21:23 AM

Quant Time: Apr 30 06:18:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	311115	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	469922	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	405255	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	158365	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	181265	45.35	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 90.70%			
35) Dibromofluoromethane	7.513	113	164357	56.66	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 113.32%			
50) Toluene-d8	10.029	98	601858	52.86	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 105.72%			
62) 4-Bromofluorobenzene	12.352	95	165369	42.93	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 85.86%			
Target Compounds						
27) cis-1,2-Dichloroethene	6.749	96	11427m	2.95	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
Data File : VN066885.D
Acq On : 30 Apr 2021 1:17
Operator : JC/MD
Sample : M2137-16
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-9-PS

Manual Integrations
APPROVED

MMDadoda
4/30/2021 11:21:23 AM

Quant Time: Apr 30 06:18:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

