

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
 Data File : VN066887.D
 Acq On : 30 Apr 2021 2:05
 Operator : JC/MD
 Sample : M2137-20
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-MW-3S(ONSITE)

Manual Integrations
 APPROVED

SAM
 5/4/2021 5:52:53 PM

Quant Time: Apr 30 06:19:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	324325	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	495639	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	428737	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	165389	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	192626	46.23	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 92.46%			
35) Dibromofluoromethane	7.513	113	176827	57.80	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 115.60%			
50) Toluene-d8	10.032	98	645019	53.71	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 107.42%			
62) 4-Bromofluorobenzene	12.349	95	178779	44.00	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 88.00%			
Target Compounds						
44) Trichloroethene	8.776	130	1010m	0.29	ug/l	Qvalue
64) Tetrachloroethene	10.573	164	3196	1.13	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
Data File : VN066887.D
Acq On : 30 Apr 2021 2:05
Operator : JC/MD
Sample : M2137-20
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-MW-3S(ONSITE)

Manual Integrations
APPROVED

SAM
5/4/2021 5:52:53 PM

Quant Time: Apr 30 06:19:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

