

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
 Data File : VN066888.D
 Acq On : 30 Apr 2021 2:29
 Operator : JC/MD
 Sample : M2137-21
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-MW-1D(OFFSITE)

Quant Time: Apr 30 06:32:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	318356	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	474401	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	407900	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	155857	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	183967	44.98	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 89.96%			
35) Dibromofluoromethane	7.513	113	166743	56.94	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 113.88%			
50) Toluene-d8	10.029	98	612651	53.30	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 106.60%			
62) 4-Bromofluorobenzene	12.349	95	142560	36.66	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 73.32%			
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	6.746	96	61929	15.62	ug/l	88
44) Trichloroethene	8.771	130	17633m	5.30	ug/l	
64) Tetrachloroethene	10.573	164	47694	17.68	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
Data File : VN066888.D
Acq On : 30 Apr 2021 2:29
Operator : JC/MD
Sample : M2137-21
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-MW-1D(OFFSITE)

Quant Time: Apr 30 06:32:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

