

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
 Data File : VN066895.D
 Acq On : 30 Apr 2021 5:17
 Operator : JC/MD
 Sample : PB136054ZHE#08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136054ZHE#08

Quant Time: Apr 30 06:34:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.594	168	318906	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	486423	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	456266	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	191425	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	188037	45.89	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.78%
35) Dibromofluoromethane	7.513	113	170854	56.90	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	113.80%
50) Toluene-d8	10.029	98	655633	55.63	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	111.26%
62) 4-Bromofluorobenzene	12.349	95	204580	51.30	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.60%
Target Compounds						
					Qvalue	
16) Acetone	3.758	43	18822	12.18	ug/l	96
20) Methylene Chloride	4.472	84	28402	7.25	ug/l	93
43) Isopropyl Acetate	8.095	43	22330	2.93	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
Data File : VN066895.D
Acq On : 30 Apr 2021 5:17
Operator : JC/MD
Sample : PB136054ZHE#08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136054ZHE#08

Quant Time: Apr 30 06:34:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

