

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
 Data File : VN066853.D
 Acq On : 29 Apr 2021 9:58
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 30 06:13:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.542	0.490	9.6	82	0.00
3 P	Chloromethane	0.711	0.641	9.8	86	0.00
4 C	Vinyl Chloride	0.695	0.655	5.8#	86	0.00
5 T	Bromomethane	0.453	0.414	8.6	86	0.00
6 T	Chloroethane	0.432	0.398	7.9	84	0.00
7 T	Trichlorofluoromethane	0.835	0.781	6.5	87	0.00
8 T	Diethyl Ether	0.339	0.298	12.1	79	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.494	0.475	3.8	89	0.00
10 T	Methyl Iodide	0.755	0.720	4.6	86	0.00
11 T	Tert butyl alcohol	0.099	0.076	23.2	71	0.00
12 CM	1,1-Dichloroethene	0.500	0.460	8.0#	83	0.00
13 T	Acrolein	0.083	0.075	9.6	84	0.00
14 T	Allyl chloride	0.940	0.772	17.9	74	0.00
15 T	Acrylonitrile	0.263	0.239	9.1	77	0.00
16 T	Acetone	0.242	0.245	-1.2	97	0.00
17 T	Carbon Disulfide	1.535	1.385	9.8	83	0.00
18 T	Methyl Acetate	0.724	0.521	28.0#	75	0.00
19 T	Methyl tert-butyl Ether	1.714	1.480	13.7	78	0.00
20 T	Methylene Chloride	0.614	0.555	9.6	85	0.00
21 T	trans-1,2-Dichloroethene	0.538	0.515	4.3	86	0.00
22 T	Diisopropyl ether	1.875	1.569	16.3	74	0.00
23 T	Vinyl Acetate	1.560	1.280	17.9	71	0.00
24 P	1,1-Dichloroethane	1.039	0.932	10.3	81	0.00
25 T	2-Butanone	0.341	0.304	10.9	77	0.00
26 T	2,2-Dichloropropane	0.882	0.793	10.1	83	0.00
27 T	cis-1,2-Dichloroethene	0.623	0.584	6.3	84	0.00
28 T	Bromochloromethane	0.477	0.399	16.4	78	0.00
29 T	Tetrahydrofuran	0.233	0.191	18.0	71	0.00
30 C	Chloroform	1.008	0.921	8.6#	83	0.00
31 T	Cyclohexane	0.993	0.820	17.4	78	0.00
32 T	1,1,1-Trichloroethane	0.862	0.777	9.9	82	0.00
33 S	1,2-Dichloroethane-d4	0.642	0.514	19.9	77	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	82	0.00
35 S	Dibromofluoromethane	0.309	0.328	-6.1	87	0.00
36 T	1,1-Dichloropropene	0.465	0.490	-5.4	83	0.00
37 T	Ethyl Acetate	0.493	0.432	12.4	69	0.00
38 T	Carbon Tetrachloride	0.446	0.471	-5.6	83	0.00
39 T	Methylcyclohexane	0.550	0.596	-8.4	83	0.00
40 TM	Benzene	1.442	1.549	-7.4	84	0.00
41 T	Methacrylonitrile	0.224	0.191	14.7	62	0.00
42 TM	1,2-Dichloroethane	0.487	0.463	4.9	74	0.00
43 T	Isopropyl Acetate	0.784	0.703	10.3	69	0.00
44 TM	Trichloroethene	0.351	0.395	-12.5	88	0.00
45 C	1,2-Dichloropropane	0.385	0.407	-5.7#	83	0.00
46 T	Dibromomethane	0.235	0.255	-8.5	83	0.00
47 T	Bromodichloromethane	0.493	0.521	-5.7	82	0.00
48 T	Methyl methacrylate	0.342	0.317	7.3	67	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
 Data File : VN066853.D
 Acq On : 29 Apr 2021 9:58
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 30 06:13:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.005	0.006	-20.0	87	0.00
50 S	Toluene-d8	1.211	1.265	-4.5	87	0.00
51 T	4-Methyl-2-Pentanone	0.426	0.415	2.6	70	0.00
52 CM	Toluene	0.867	0.954	-10.0#	85	0.00
53 T	t-1,3-Dichloropropene	0.527	0.561	-6.5	79	0.00
54 T	cis-1,3-Dichloropropene	0.591	0.638	-8.0	82	0.00
55 T	1,1,2-Trichloroethane	0.333	0.377	-13.2	86	0.00
56 T	Ethyl methacrylate	0.453	0.495	-9.3	76	0.00
57 T	1,3-Dichloropropane	0.570	0.612	-7.4	82	0.00
58 T	2-Chloroethyl Vinyl ether	0.102	0.102	0.0	72	0.00
59 T	2-Hexanone	0.253	0.283	-11.9	72	0.00
60 T	Dibromochloromethane	0.363	0.420	-15.7	87	0.00
61 T	1,2-Dibromoethane	0.334	0.380	-13.8	84	0.00
62 S	4-Bromofluorobenzene	0.410	0.450	-9.8	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.331	0.376	-13.6	96	0.00
65 PM	Chlorobenzene	1.006	1.079	-7.3	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.375	0.411	-9.6	89	0.00
67 C	Ethyl Benzene	1.764	1.888	-7.0#	85	0.00
68 T	m/p-Xylenes	0.647	0.736	-13.8	88	0.00
69 T	o-Xylene	0.646	0.701	-8.5	86	0.00
70 T	Styrene	0.976	1.153	-18.1	87	0.00
71 P	Bromoform	0.272	0.325	-19.5	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	3.665	3.592	2.0	86	0.00
74 T	N-amyl acetate	0.866	0.715	17.4	64	0.00
75 P	1,1,2,2-Tetrachloroethane	1.206	1.135	5.9	85	0.00
76 T	1,2,3-Trichloropropane	1.054	0.902	14.4	72	0.00
77 T	Bromobenzene	0.904	0.937	-3.7	92	0.00
78 T	n-propylbenzene	4.042	4.143	-2.5	87	0.00
79 T	2-Chlorotoluene	2.571	2.487	3.3	88	0.00
80 T	1,3,5-Trimethylbenzene	3.094	3.093	0.0	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.314	0.302	3.8	76	0.00
82 T	4-Chlorotoluene	2.481	2.556	-3.0	90	0.00
83 T	tert-Butylbenzene	2.651	2.614	1.4	90	0.00
84 T	1,2,4-Trimethylbenzene	2.914	2.986	-2.5	88	0.00
85 T	sec-Butylbenzene	3.438	3.521	-2.4	89	0.00
86 T	p-Isopropyltoluene	2.951	3.149	-6.7	91	0.00
87 T	1,3-Dichlorobenzene	1.537	1.645	-7.0	94	0.00
88 T	1,4-Dichlorobenzene	1.634	1.679	-2.8	98	0.00
89 T	n-Butylbenzene	2.515	2.453	2.5	84	0.00
90 T	Hexachloroethane	0.618	0.617	0.2	93	0.00
91 T	1,2-Dichlorobenzene	1.558	1.635	-4.9	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.216	0.177	18.1	78	0.00
93 T	1,2,4-Trichlorobenzene	0.863	0.954	-10.5	95	0.00
94 T	Hexachlorobutadiene	0.479	0.547	-14.2	107	0.00
95 T	Naphthalene	2.221	2.107	5.1	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042921\
Data File : VN066853.D
Acq On : 29 Apr 2021 9:58
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Apr 30 06:13:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.834	0.900	-7.9	96	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6