

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042922\
 Data File : VN072262.D
 Acq On : 29 Apr 2022 12:00
 Operator : JC\MD
 Sample : VX0429WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VX0429WBL01

Quant Time: Apr 30 03:11:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	205277	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	340807	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	295485	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	90393	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	173262	54.978	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.960%	
35) Dibromofluoromethane	8.022	113	112412	51.533	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.060%	
50) Toluene-d8	10.439	98	415751	48.504	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 97.000%	
62) 4-Bromofluorobenzene	12.727	95	123558	43.430	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 86.860%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042922\
Data File : VN072262.D
Acq On : 29 Apr 2022 12:00
Operator : JC\MD
Sample : VX0429WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VX0429WBL01

Quant Time: Apr 30 03:11:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

