

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042922\
 Data File : VN072266.D
 Acq On : 29 Apr 2022 13:46
 Operator : JC\MD
 Sample : N2617-11DL 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-14DL

Quant Time: Apr 30 03:13:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	196672	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	332352	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	288688	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	97679	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	172194	57.030	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 114.060%			
35) Dibromofluoromethane	8.016	113	111117	52.235	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 104.480%			
50) Toluene-d8	10.439	98	410702	49.134	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 98.260%			
62) 4-Bromofluorobenzene	12.727	95	132808	47.868	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 95.740%			
Target Compounds					Qvalue	
19) Methyl tert-butyl Ether	5.622	73	19038	2.586	ug/l	# 1
31) Cyclohexane	8.086	56	14887	3.690	ug/l	# 76
39) Methylcyclohexane	9.451	83	5356	1.547	ug/l	# 88
40) Benzene	8.457	78	85509	8.903	ug/l	99
52) Toluene	10.498	92	16193	2.820	ug/l	98
67) Ethyl Benzene	11.839	91	491841	48.350	ug/l	97
68) m/p-Xylenes	11.951	106	49755	13.172	ug/l	85
69) o-Xylene	12.274	106	14701	4.014	ug/l	92
73) Isopropylbenzene	12.574	105	17988	2.085	ug/l	98
78) n-propylbenzene	12.915	91	48991	5.155	ug/l	94
80) 1,3,5-Trimethylbenzene	13.057	105	9620	1.339	ug/l	97
84) 1,2,4-Trimethylbenzene	13.363	105	129128	18.576	ug/l	95
95) Naphthalene	15.504	128	12340	6.529	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042922\
Data File : VN072266.D
Acq On : 29 Apr 2022 13:46
Operator : JC\MD
Sample : N2617-11DL 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-14DL

Quant Time: Apr 30 03:13:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

