

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042922\
 Data File : VN072267.D
 Acq On : 29 Apr 2022 14:10
 Operator : JC\MD
 Sample : N2617-07DL 40X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-8IDL

Quant Time: Apr 30 03:13:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	201898	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	346064	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	301053	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	105334	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	184662	59.576	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 119.160%	
35) Dibromofluoromethane	8.016	113	120838	54.554	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 109.100%	
50) Toluene-d8	10.439	98	449212	51.612	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.220%	
62) 4-Bromofluorobenzene	12.727	95	145940	50.517	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.040%	
Target Compounds						
					Qvalue	
16) Acetone	4.298	43	3553	2.659	ug/l	96
20) Methylene Chloride	5.098	84	5685	2.219	ug/l #	92
31) Cyclohexane	8.086	56	22638	5.466	ug/l #	79
39) Methylcyclohexane	9.451	83	7785	2.159	ug/l	97
40) Benzene	8.457	78	82248	8.224	ug/l	98
52) Toluene	10.504	92	64905	10.856	ug/l	98
67) Ethyl Benzene	11.839	91	752639	70.949	ug/l	96
68) m/p-Xylenes	11.945	106	213631	54.232	ug/l	87
69) o-Xylene	12.274	106	48555	12.715	ug/l	92
73) Isopropylbenzene	12.574	105	24135	2.594	ug/l	99
78) n-propylbenzene	12.916	91	71816	7.007	ug/l	97
80) 1,3,5-Trimethylbenzene	13.051	105	21043	2.716	ug/l	99
84) 1,2,4-Trimethylbenzene	13.363	105	101564	13.549	ug/l	95
95) Naphthalene	15.504	128	52668	13.123	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042922\
Data File : VN072267.D
Acq On : 29 Apr 2022 14:10
Operator : JC\MD
Sample : N2617-07DL 40X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-8IDL

Quant Time: Apr 30 03:13:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

