

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042922\
 Data File : VN072271.D
 Acq On : 29 Apr 2022 15:46
 Operator : JC\MD
 Sample : N2617-20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-2022-04-28

Quant Time: Apr 30 03:14:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	185195	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	315203	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	265669	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	78904	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	173278	60.946	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 121.900%	
35) Dibromofluoromethane	8.016	113	109846	54.447	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 108.900%	
50) Toluene-d8	10.439	98	403994	50.961	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.920%	
62) 4-Bromofluorobenzene	12.727	95	115900	44.047	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 88.100%	
Target Compounds						
16) Acetone	4.299	43	7412	6.047	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042922\
Data File : VN072271.D
Acq On : 29 Apr 2022 15:46
Operator : JC\MD
Sample : N2617-20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-2022-04-28

Quant Time: Apr 30 03:14:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

