

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
 Data File : VN077516.D
 Acq On : 29 Apr 2023 14:03
 Operator : JC\MD
 Sample : PB152434TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152434TB

Quant Time: May 01 01:23:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	569242	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1077434	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	1014535	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	344881	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.578	65	378164	50.911	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.820%
35) Dibromofluoromethane	8.172	113	332290	51.312	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.620%
50) Toluene-d8	10.572	98	1343222	54.120	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.240%
62) 4-Bromofluorobenzene	12.854	95	425433	52.392	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.780%
Target Compounds						
						Qvalue
16) Acetone	4.443	43	44282	18.739	ug/l #	87
20) Methylene Chloride	5.284	84	27060	3.611	ug/l #	86
25) 2-Butanone	7.495	43	32613	9.730	ug/l	95
37) Ethyl Acetate	7.566	43	95075	13.373	ug/l	98
43) Isopropyl Acetate	8.695	43	404720	32.907	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
Data File : VN077516.D
Acq On : 29 Apr 2023 14:03
Operator : JC\MD
Sample : PB152434TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152434TB

Quant Time: May 01 01:23:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

