

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
 Data File : VN077517.D
 Acq On : 29 Apr 2023 14:27
 Operator : JC\MD
 Sample : PB152434ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152434ZHE#01

Quant Time: May 01 01:23:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	624064	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1188744	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	1113638	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	382402	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	417592	51.280	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.560%
35) Dibromofluoromethane	8.172	113	364162	50.968	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.940%
50) Toluene-d8	10.566	98	1485614	54.253	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.500%
62) 4-Bromofluorobenzene	12.854	95	478286	53.385	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.780%
Target Compounds						
16) Acetone	4.443	43	85705	33.082	ug/l	Qvalue 99
20) Methylene Chloride	5.290	84	42009	5.113	ug/l #	95
37) Ethyl Acetate	7.572	43	104129	13.275	ug/l	99
43) Isopropyl Acetate	8.695	43	488146	35.973	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
Data File : VN077517.D
Acq On : 29 Apr 2023 14:27
Operator : JC\MD
Sample : PB152434ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152434ZHE#01

Quant Time: May 01 01:23:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

