

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
 Data File : VN077519.D
 Acq On : 29 Apr 2023 15:15
 Operator : JC\MD
 Sample : PB152434ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152434ZHE#03

Quant Time: May 01 01:23:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	574219	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1120101	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	1061830	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	377909	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	401687	53.609	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.220%	
35) Dibromofluoromethane	8.172	113	341717	50.758	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.520%	
50) Toluene-d8	10.571	98	1383480	53.619	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.240%	
62) 4-Bromofluorobenzene	12.854	95	457213	54.161	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 108.320%	
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	85606	35.912	ug/l	98
20) Methylene Chloride	5.289	84	38815	5.134	ug/l	86
37) Ethyl Acetate	7.566	43	102494	13.868	ug/l	99
43) Isopropyl Acetate	8.695	43	529298	41.396	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
Data File : VN077519.D
Acq On : 29 Apr 2023 15:15
Operator : JC\MD
Sample : PB152434ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152434ZHE#03

Quant Time: May 01 01:23:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

