

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
 Data File : VN077520.D
 Acq On : 29 Apr 2023 15:40
 Operator : JC\MD
 Sample : PB152434ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152434ZHE#04

Quant Time: May 01 01:24:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	661498	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1281838	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	1157924	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	412369	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	446001	51.670	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.340%	
35) Dibromofluoromethane	8.172	113	399188	51.813	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.620%	
50) Toluene-d8	10.572	98	1550517	52.510	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.020%	
62) 4-Bromofluorobenzene	12.854	95	503098	52.076	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.160%	
Target Compounds						
					Qvalue	
16) Acetone	4.437	43	86084	31.347	ug/l	95
20) Methylene Chloride	5.296	84	40105	4.605	ug/l	96
37) Ethyl Acetate	7.566	43	105785	12.507	ug/l	97
43) Isopropyl Acetate	8.695	43	577683	39.480	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
Data File : VN077520.D
Acq On : 29 Apr 2023 15:40
Operator : JC\MD
Sample : PB152434ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152434ZHE#04

Quant Time: May 01 01:24:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

